



UT39E QA测试报告

使用设备: FLUKE5522A

环境温湿度: 23.5°C/56.2%

功能	量程	输入值	准确度	误差范围	测试值	测试结果
DCV	200mV	190.00mV	±(0.5%+1)	189.87~190.13mV	190.01mV	合格
	2V	1.9000V		1.8978~1.9022V	1.9002V	合格
	20V	19.000V		18.978~19.022V	19.001V	合格
	200V	190.00V		189.78~190.22V	189.99V	合格
	1000V	1000.0V	±(0.8%+2)	998.0~1002.0V	1000.2V	合格
ACV	2V	1.9000V/50Hz	±(0.5%+10)	1.8895~1.9105V	1.8999V	合格
	20V	19.000V/400Hz		18.895~19.105V	18.998V	合格
	200V	190.00V/60Hz		188.95~191.05V	190.01V	合格
	750V	750.0V/400Hz	±(0.8%+15)	742.5~757.5V	750.0V	合格
DCA	2mA	1.9000mA	±(0.5%+5)	1.8900~1.9100mA	1.8996mA	合格
	200mA	190.00mA	±(0.8%+5)	188.43~191.57mA	190.00mA	合格
	20A	10.000A	±(2%+10)	9.790~10.210A	10.002A	合格
ACA	2mA	1.9000mA/50Hz	±(0.8%+10)	1.8838~1.9162mA	1.9010mA	合格
	200mA	190.00mA/60Hz	±(1.2%+10)	187.62~192.38mA	189.96mA	合格
	20A	10.000A/400Hz	±(2.5%+10)	9.740~10.260A	10.006A	合格
Ω	200Ω	190.00Ω	±(0.5%+10)	188.95~191.05Ω	190.00Ω	合格
	2kΩ	1.9000kΩ	±(0.3%+1)	1.8942~1.9058kΩ	1.9001kΩ	合格
	20kΩ	19.000kΩ		18.942~19.058kΩ	18.996kΩ	合格
	2MΩ	1.9000MΩ		1.8942~1.9058MΩ	1.8996MΩ	合格
	200MΩ	190.00MΩ	±[5%(读数-100)+10]	180.40~199.60MΩ	189.92MΩ	合格
CAP	2nF	1.9000nF	±(4%+10)	1.8230~1.9770nF	1.9010nF	合格
	20nF	19.000nF		18.230~19.770nF	19.018nF	合格
	200nF	190.00nF		182.30~197.70nF	190.20nF	合格
	20μ F	19.000μ F		18.230~19.770uF	18.984uF	合格
HZ	2kHz	1.9000kHz/200mVrms	±(2%+5)	1.8642~1.9358KHZ	1.9005KHZ	合格
	20kHz	19.000kHz/200mVrms	±(1.5%+5)	18.710~19.290KHZ	19.002KHZ	合格
二极管		IN4007	/	0.5~0.8V	0.602V	合格
通断		10Ω	/	发声		合格



检测员: 张琼

审核员: 郭桂余