

UT256 QA测试报告

检测仪器: FLUKE5522A

环境温湿度: 24.1°C/50.2%

功能	量程	分辨率	精度±(a%读数+b 字数)		输入	单位	压缩	读数范围			测试值	结果	
			±a%	b									
交流电流 ACA	60.0A	0.1	±	2.5%	10	1.0	A(50Hz) 180°	90%	0.1	~	1.9	1.0	合格
		0.1	±	2.5%	10	3.0	A(50Hz) 180°	90%	2.0	~	4.0	2.9	合格
		0.1	±	2.5%	5	10.0	A(50Hz) 180°	80%	9.4	~	10.6	9.9	合格
		0.1	±	2.5%	5	30.0	A(50Hz) 180°	80%	29.0	~	31.0	29.9	合格
		0.1	±	2.5%	5	30.0	A(60Hz) 180°	80%	29.0	~	31.0	29.9	合格
		0.1	±	2.5%	5	50.0	A(50Hz) 0°	80%	48.6	~	51.4	49.9	合格
		0.1	±	2.5%	5	50.0	A(50Hz) 90°	80%	48.6	~	51.4	50.0	合格
		0.1	±	2.5%	5	50.0	A(50Hz) 180°	80%	48.6	~	51.4	50.0	合格
		0.1	±	2.5%	5	50.0	A(60Hz) 180°	80%	48.6	~	51.4	50.1	合格
直流电流 DCA	60.0A	0.1	±	2.0%	10	1.0	180°	90%	0.1	~	1.9	0.9	合格
		0.1	±	2.0%	10	3.0	180°	90%	2.0	~	4.0	2.9	合格
		0.1	±	2.0%	5	10.0	180°	80%	9.4	~	10.6	9.9	合格
		0.1	±	2.0%	5	30.0	180°	80%	29.1	~	30.9	29.9	合格
		0.1	±	2.0%	5	50.0	0°	80%	48.8	~	51.2	50.2	合格
		0.1	±	2.0%	5	50.0	90°	80%	48.8	~	51.2	50.2	合格
		0.1	±	2.0%	5	50.0	180°	80%	48.8	~	51.2	50.2	合格
		0.1	±	2.0%	5	50.0	270°	80%	48.8	~	51.2	50.3	合格

检测员: 黄满妮

审核员: 黄满妮

