

测试报告

仪器型号: LM560G

检验项目及要求		结果						备注		
外壳及支架		OK			OK					
安平±4°		OK			OK					
开关运转状况		OK			OK					
各控制键及开关		OK			OK					
调制状态		OK			OK					
H		左	中	右		左	中	右		
	直线度 <±2mm/7m	H1	0.3 mm	0.3 mm	0.3 mm		0.3 mm	0.3 mm	0.3 mm	
		H2	0.3 mm	0.3 mm	0.3 mm		0.3 mm	0.3 mm	0.3 mm	
		H3	\	\	\		\	\	\	
		H4	\	\	\		\	\	\	
精度:<±2mm/7m	±0.5mm			±0.5mm						
激光线质		OK			OK					

检验项目及要求		结果								备注	
天顶:<±1mm/3m											
V		V1	V2	V3	V4	V1	V2	V3	V4		
	<±2mm/7m	上	0.3 mm	0.3 mm	0.3 mm	0.3 mm	0.3 mm	0.3 mm	0.3 mm	0.3 mm	
		中	0.3 mm	0.3 mm	0.3 mm	0.3 mm	0.3 mm	0.3 mm	0.3 mm	0.3 mm	
		下	0.3 mm	0.3 mm	0.3 mm	0.3 mm	0.3 mm	0.3 mm	0.3 mm	0.3 mm	
	V1-V3, V2-V4: 90°	OK	OK	OK	OK	OK	OK	OK	OK		
V1-V3, V2-V4: 180°	OK	OK	OK	OK	OK	OK	OK	OK			
下对点:<±2mm/2mm		0.1mm									
其它		\									
结论:		合格									

优利德科技(中国)股份有限公司
校准专用章

首检员: 刘爱

审核员: 郭桂余